



COMPARATIVE CHART OF ULTRASONIC COUPLANTS CHARACTERISTICS

	APPEARANCE	ODOR	DENSITY AT 20°C	pH	WASHABILITY	RECOMMENDED DILUTION IN POTABLE WATER	PEELING
GREEN GEL U 45	Transparent green gel	Odorless	1,03 Kg/l	8	Water	From 0 to 1:1	No
COUPLANT U 47	Transparent gel	Odorless	1,00 Kg/l	7-8	Water	From 0 to 1:1	No
COUPLANT U 49	Transparent gel	Odorless	1,00 Kg/l	7-8	Water	From 0 to 1:1	Yes
COUPLANT U 51	White gel	Odorless	1,03 Kg/l	7-8	Water	From 0 to 1:1	No
COUPLANT U 53	White gel	Odorless	1,03 Kg/l	7-8	Water	From 0 to 1:1	Yes
POWDER U 55+AC	Powder to prepare transparent gel	Odorless	NA	~7	Water	33 g/l	NA

Typical values.